

Product Safety and Quality
EMC and Telecommunications Services

October 15, 2012

WOO Yim Leng (Ms)
Senior Manager, Resource Management & Standards Division
Infocomm Development Authority of Singapore
10 Pasir Panjang Road
#10-01 Mapletree Business City
Singapore 117438

To Whom It May Concern:

Sub: New IDA Technical /Reference specifications and documents for public comments

We TÜV Rheinland of North America are in favor of the change and support of the Singapore iDA standard revision of the following

- a) Allow the use of 9 -315Khz for medical and biological telemetry with field strength up to 30dBuA/m @10m
- b) Revision of the field strength of induction loop devices in band 16 -150Hkz from 66dBuA/m@ 3m to 66dBuA/m@10m.

Respectfully,



Suresh Kondapalli
Senior Wireless Engineer
TCB and MRA program Manager
San Francisco Region

TUV Rheinland
of North America, Inc.
San Francisco Office

1279 Quarry Lane
Suite A
Pleasanton, CA 94566

Tel 925-249-9123
Fax 925-249-9124
Mail info@us.tuv.com

North American Headquarters

12 Commerce Road
Newtown, CT 06470

Tel 203-426-0888
Fax 203-426-4009
Toll Free TÜV-RHEINLAND
Web www.tuv.com

Member of
TÜV Rheinland Group